

CERTIFICATE
of Conformity



Registration No.: AK 50108575 0001

Report No.: 10019829 001

Holder: Radiant Innovation Inc.
1F, No. 3, Industrial E. 9th Rd.
Science-Based Industrial Park
HsinChu 300
Taiwan

Product: Thermometer
(IR Thermometer)

Identification: TH6XYZ
(X = 0~9, A~Z, or blank; Y = F or N; Z = E or blank)

Tested acc. to: EN 60601-1:1990+A1+A2+A13
IEC 60601-1:1988+A1+A2

The certificate of conformity refers to the above mentioned product. This is to certify that the specimen is in conformity with the assessment requirement mentioned above. This certificate does not imply assessment of the production of the product and does not permit the use of a TÜV Rheinland mark of conformity.

Cologne, 20.06.2007



Certification Body


Dipl.-Ing. F. Stoelzel

TÜV Rheinland Product Safety GmbH - Am Grauen Stein - D-51105 Köln